




CERTIFICATE OF ANALYSIS

PRODUCT: CAPILLARY PERFORMANCE RUN BUFFER A
PART NO.: 338426 LOT NO.: M910053
EXPIRATION DATE: 2021-10-14

THIS REAGENT WAS TESTED AND RELEASED BY BECKMAN COULTER QUALITY CONTROL LABORATORIES AND THE FOLLOWING RESULTS WERE OBTAINED.

TEST	SPECIFICATION	TEST RESULTS
pH	8.35 ± 0.05	PASS
MIGRATION TIME OF FIRST PEAK	4.75 ± 0.75 MINUTES	PASS
% RSD OF MIGRATION TIME OF EACH PEAK	≤ 2 %	PASS


STEPHANIE YAN
QUALITY CONTROL

24 Mar 2020
March 24, 2020
DATE